

HEARING TECHNOLOGY



Arne Leijon
Professor of Hearing Technology

The long-term goal of research in Hearing Technology is to understand

1. how the physiological state of the human auditory system affects the ability to detect and analyse sound,
2. how an impaired ability to detect and analyse sound affects the ability to understand and produce speech,
3. how personal aids should be designed and individually fitted to compensate for hearing impairment.

The group pursues several applied projects concerning the design, fitting, and evaluation of personal aids for the hearing impaired, as well as methods and systems for teaching and training of speech perception and speech production.

Highlights 2001

Two major milestones were reached during the year 2001:

1. A EU project (SPECO) was successfully concluded, resulting in a new audio-visual pronunciation teaching and training method and software system for speech- and hearing-impaired children.
2. A new theoretical approach was proposed to the problem of estimating the capacity of the auditory system to convey speech information to the brain. This capacity sets an ultimate limit for speech recognition. The method can be extended to auditory-visual integration and to the electrical information transmission in cochlear implants.

Signal processing algorithms for hearing instruments

Advanced hearing instruments must adapt their characteristics automatically to varying acoustic listening conditions. A new algorithm for this purpose has been developed and evaluated in an industry-related project. The algorithm is based on hidden Markov models. The method is robust, as it uses only modulation properties of the signal and requires no information about absolute signal levels or spectral densities. It has been evaluated for a great variety of speech environments, with various signal-to-noise ratios, and was shown to correctly classify sound environments consisting of speech in various types of background noise. Classification of music environments still remains an unsolved problem for future study.

Work has been continued on the acoustic measurement and analysis of advanced adaptive hearing instruments. A new HMM-based approach has been used to illustrate the effects of fast syllabic compression using real speech as test signals.

Individual fitting of advanced hearing instruments

Several methods for computer-aided fitting of advanced (digital) hearing aids have been studied. The results led to a need for further study on the psycho-acoustical assumptions underlying most of these methods. This work was pursued in a series of experiments at the National Acoustic Laboratories in Sydney. Preliminary results cast some doubt on the conventional assumption that hearing aids should restore normal loudness for speech, at least if conventional loudness models are used.

A special method for hearing-aid fitting, based on individual tests of masked short-tone thresholds, has been developed and tested at the hearing clinic at the Karolinska hospital in Stockholm. We have co-operated with this work in order to evaluate some of the effects of basic auditory-signal analysis on speech recognition.

Multi-modal sensory aids

Together with researchers in the speech group we are investigating a new method to support audio-visual telephone communication for hearing-impaired persons without special equipment for transmitting video information. The aim of this work is to generate an artificial talking face which can be used as a lip-reading support by the hearing-impaired telephone user. The articulation movements of the artificial face are controlled by parameters automatically derived from the speech signal. Tests with hearing-impaired listeners have demonstrated that the artificial face can improve the intelligibility of telephone speech, even when the artificial face is controlled by an imperfect recognition system. Current work attempts to increase the benefit even further by improving the recognition system.

Evaluation of cochlear implants

The work on comparing results from conventional hearing aids and cochlear implants (CI) has been continued in close cooperation with the CI clinic at Huddinge hospital. The test battery includes measurement of time resolution, auditory, visual and audio-visual perception of consonants, vowels, prosodic features and speech tracking, as well as subjective assessment. The aim of the study is to obtain data to support criteria for patient selection for cochlear implant operations, and also to evaluate the relative importance of various speech cues for speech-reading support.

Teaching and training systems

We participate in a EU project (Ortho-Logo-Paedia, OLP), which aims at improving the quality of life of persons with articulatory impairments by applying a novel technological aid to speech therapy, by integrating this training with speech recognition technology and by making these facilities available over the Internet. Four basic types of pathologies will be addressed in the OLP project. These are dysarthria, craniofacial disorders, functional articulation disorders, and hearing impairments.

Filename: hearing2.doc
Directory: I:\annual\Annual2001
Template: J:\tmh.kth.se\tmh\proj\common\template\tmh\Normal.dot
Title: A look at 2001
Subject:
Author: andersa
Keywords:
Comments:
Creation Date: 2002-04-25 08:51
Change Number: 10
Last Saved On: 2002-05-23 09:39
Last Saved By: Personal
Total Editing Time: 43 Minutes
Last Printed On: 2002-06-12 10:43
As of Last Complete Printing
Number of Pages: 2
Number of Words: 742 (approx.)
Number of Characters: 4 235 (approx.)